



File
LR1409

CERTIFICATE OF COMPLIANCE (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Casella Holdings Limited
Address	Regent House, Wolseley Road Kempston, Bedford Bedfordshire MK42 7JY United Kingdom
Project Number	LR1409-2R5
Product	Personal Air Sampling Pump
Model Number	APEX2
Electrical Ratings	See ANNEX below
Markings	See ANNEX below
Applicable Standards	CSA C22.2 No. 60079-0:2015 CSA C22.2 No. 60079-11:2014 CSA C22.2 No. 61010-1:2012 UL 60079-0 6th ed UL 60079-11 6th ed UL 913 8th ed UL 61010-1 3rd ed
Factory/Manufacturing Location	Casella Holdings Limited Regent House, Wolseley Road, Kempston, Bedford Bedfordshire MK42 7JY, United Kingdom
Conditions of Certification	<p>The Casella APEX2 Personal Air Sampling Pump is a personal air sampling pump typically used to take samples of airborne materials or toxic contaminants. It is a bodily worn, battery-powered self-contained device housed within a PC anti-static enclosure. The pump is designed to provide a stable controlled flow rate of approximately 0.5 to 5 litre/min such that a known volume of air is passed through a passive sampling head and filter medium.</p> <p>Airborne materials are collected onto a filter substrate and are subjected to mass or chemical analysis in a laboratory in order to establish a worker's exposure to potentially hazardous materials. Sampling pumps are typically worn on the user's belt and are connected to the sampling head via a short length of tube.</p> <p>The equipment may only be recharged in the safe area using either:</p> <ol style="list-style-type: none"> 1. APEX2/IS 5L Pump Five Way Charger 2. APEX2/IS 5L Pump Single Way Charger <p>The chargers must be supplied from an SELV source with $U_m = 63V$</p> <p>These have an output of:</p> <p>$U_m = 14.5V$ $I_{max} = 850mA$</p> <p>When connected to a USB in the safe area:</p> <p>$U_m = 5.9V$ $I_{max} = 85mA$</p>





QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
LR1409

Statement of Compliance: The product(s)/equipment identified in this Certificate and described in the Certification Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.

IMPORTANT NOTE: In order to maintain the integrity of the QPS Mark(s), certification will be revoked if:

- (1) Compliance to the above-mentioned Standard(s), or those identified in future QPS Standard Update Notice – SUN (QSD 55) is not maintained, or,
- (2) If the product/equipment is modified after certification is granted without prior written consent by QPS



Issued By: Dave Adams, P.Eng.

Signature: 

Date: September 6, 2024



The SCC and IAS Accreditation Symbols are official symbols of the respective accreditation bodies, used under license.

81 Kelfield St., Unit 8, Toronto, ON M9W 5A3 Tel: 416-241-8857; Fax: 416-241-0682

www.qps.ca





QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
LR1409

ANNEX:

Electrical ratings:

The chargers must be supplied from an SELV source with $U_m = 63V$

These have an output of:

$U_m = 14.5V$

$I_{max} = 850mA$

When connected to a USB in the safe area:

$U_m = 5.9V$

$I_{max} = 85mA$

Markings:

The APEX2 Personal Air Sampling Pump has the following classifications:

Canada	US
Ex ia I Ma Ex ia IIC T4 Ga Ex ia IIIC T135°C Da Ta = -20°C to +45°C	Class I, Zone 0 AEx ia IIC T4 Ga Zone 20 AEx ia IIIC T135°C Da



The SCC and IAS Accreditation Symbols are official symbols of the respective accreditation bodies, used under license.

81 Kelfield St., Unit 8, Toronto, ON M9W 5A3 Tel: 416-241-8857; Fax: 416-241-0682

www.qps.ca

